

# Counting Solutions

## LSC Technical Tips from PerkinElmer

---

CS-001 (01/10/95)

### Radio-Carbon Dioxide ( $^{14}\text{CO}_2$ ) Trapping and Counting

By: J. Thomson and D.A. Burns

#### Introduction

The Sample Oxidiser is perfectly designed to trap  $^{14}\text{CO}_2$  originating from the combustion of large amounts of samples containing  $^{14}\text{C}$ . However discrete gaseous  $^{14}\text{CO}_2$  samples not originating from combustion require a completely different approach for the trapping of the gas and subsequent liquid scintillation counting (LSC).  $^{14}\text{CO}_2$  gas samples originate from a variety of sources, including:

1.  $^{14}\text{CO}_2$  in expired breath
2.  $^{14}\text{CO}_2$  expired by plants
3.  $^{14}\text{CO}_2$  expulsion from blood
4.  $^{14}\text{CO}_2$  release in enzymatic studies

There are a large number of potentially useful reagents available for trapping carbon dioxide, including the following:

1. Sodium hydroxide (0.1 - 1.0 M)
2. Potassium hydroxide (0.1 - 1.0 M)
3. Hyamine hydroxide in methanol (1.0 M)
4. Ethanolamine®
5. Carbo-Sorb®
6. Carbo-Sorb E

Table 1 lists some basic information which will be helpful in selecting the best trapping reagent for a particular application.

	mM CO <sub>2</sub> per mL	mL required for 1 mM CO <sub>2</sub>	mL required for 5 mM CO <sub>2</sub>	mL required for 10 mM CO <sub>2</sub>	Flash-Point (°C)
0.1M Sodium/Potassium hydroxide	0.05	20.00	-	-	-
1.0M Sodium/Potassium hydroxide	0.50	2.00	10.00	-	-
1.0M Hyamine hydroxide in methanol	0.50	2.00	10.00	-	11°C
Ethanolamine	8.10	0.12	0.62	1.23	93°C
Carbo-Sorb	5.80	0.17	0.86	1.72	7°C
Carbo-Sorb E	4.80	0.21	1.04	2.08	27°C

**Table 1.**

Trapping capacity of suitable reagents for carbon dioxide.

Having selected a suitable carbon dioxide trapping reagent for the application, the next step is to ensure that a compatible liquid scintillation cocktail is chosen. Therefore, each reagent will be considered separately.

### **1&2. Sodium/Potassium hydroxide**

Sodium hydroxide absorbs/traps CO<sub>2</sub> by a reaction which produces a sodium carbonate solution. Potassium hydroxide performs in a similar way by forming potassium carbonate.

#### **Recommended LSC Cocktails**

1. 10 mL of Emulsifier-Safe™ #\* will accept up to 2 mL of 0.1 M sodium hydroxide/CO<sub>2</sub>.
2. 10 mL of Opti-Fluor®\* will accept up to 5 mL of 0.1 M sodium hydroxide/CO<sub>2</sub>.
3. 10 mL of Hionic-Fluor™ will accept up to 5 mL of 1.0 M sodium hydroxide fully saturated with CO<sub>2</sub>.
4. 10 mL of Ultima-Flo AF™\* will accept 10 mL of 0.5 M NaOH/CO<sub>2</sub> or 5 mL of 1.0 M NaOH/CO<sub>2</sub>.

# Sold as Poly-Fluor™ in the U.S.

\* High flash-point cocktails.

#### **Notes:**

1. Hionic-Fluor is suitable for use with hydroxide/carbonate solutions due to its sample capacity for concentrated solutions and alkaline pH. Ideally the final pH should be above 9 to avoid liberation of trapped CO<sub>2</sub>.
2. Cocktails containing mixed surfactant systems, such as Ultima Gold™ or Ultima Gold XR can be used, however counting should be performed the same day as these cocktails have the potential for slow release of CO<sub>2</sub> on prolonged storage (*i.e.* characterized by dropping CPM levels).

### **3. Hyamine hydroxide in methanol**

Chemically, Hyamine hydroxide performs similarly to sodium and potassium hydroxide in that it forms hyamine carbonate on reaction with CO<sub>2</sub>.

#### **Recommended LSC Cocktails**

1. 10 mL of Insta-Fluor™ will accept up to 7.5 mL of Hyamine hydroxide saturated with carbon dioxide.
2. 10 mL of Emulsifier-Safe™ will accept up to 3 mL of Hyamine hydroxide saturated with carbon dioxide providing a safer system due to the high flash-point of this LSC cocktail.

#### **Note:**

Foaming of Hyamine hydroxide used to absorb carbon dioxide expelled in rat breath has been reported. This can be overcome by the addition of one drop of silicone antifoam per 10 mL of Hyamine hydroxide.

The chemiluminescence resistance of both these cocktails is shown in Table 2, which clearly demonstrates their suitabilities for use with this reagent system.

	Up to 3,0 mL Hyamine hydroxide/CO <sub>2</sub> in 10 mL Emulsifier-Safe			Up to 7.5 mL Hyamine hydroxide/CO <sub>2</sub> in 10 mL Insta-Fluor			
Counting window	0 - 156	2 - 156	4 - 156	Counting window	0 - 156	2 - 156	4 - 156
CPM after one minute	53	32	24	CPM after one minute	35	28	22
CPM after five minutes	39	27	22	CPM after five minutes	34	30	27
CPM after one hour	36	31	25	CPM after one hour	24	23	18

1) All counting in a PerkinElmer Tri-Carb 1900@ 20 °C.  
 2) All samples fully saturated with carbon dioxide.  
 3) All counting in high performance glass vials.

**Table 2.**

Chemiluminescence resistance of LSC cocktails to Hyamine hydroxide/CO<sub>2</sub>.

#### **4. Ethanolamine**

Ethanolamine reacts with CO<sub>2</sub> in a rather different way than previously discussed reagents in that a carbamate is formed as opposed to a carbonate. The main difference between these two species is that a carbamate is more stable under slightly acidic conditions whereas a carbonate reacts rapidly with acids to release carbon dioxide.

Ethanolamine/CO<sub>2</sub> is notoriously difficult to incorporate into LSC cocktails and consequently the recommended LSC solutions may seem a little unusual.

#### **Recommended LSC Cocktails**

The main system known which is capable of accepting this unusual reagent is shown in Table 3. The recommended LSC cocktail requires the use of a co-solvent *e.g.* methyl cellosolve to facilitate the take up of the ethanolamine/CO<sub>2</sub>.

<b>COCKTAIL</b>	<b>Cocktail Volume (mL)</b>	<b>Methyl Cellosolve Volume (mL)</b>	<b>Ethanolamine Volume (mL)</b>	<b>mM CO<sub>2</sub> Trapping Capacity</b>
<b>Hionic-Fluor</b>	<b>10.0</b>	<b>4.0</b>	<b>1.0</b>	<b>8.1</b>
<b>Hionic-Fluor</b>	<b>10.0</b>	<b>6.5</b>	<b>2.0</b>	<b>16.2</b>
<b>Hionic-Fluor</b>	<b>10.0</b>	<b>8.5</b>	<b>3.0</b>	<b>24.3</b>

***Table 3.***

Suitable LSC solutions and capacity for ethanolamine/CO<sub>2</sub>.

#### **5&6. Carbo-Sorb/Carbo-Sorb E**

These amines react with CO<sub>2</sub> to form a carbamate as well. Both reagents were developed to work in the PerkinElmer Sample Oxidizer. It is however, possible to use these reagents as carbon dioxide trapping agents outside of the Oxidizer. As with the Oxidizer, the recommended cocktail is Permafluor® E<sup>+</sup>. When using Carbo-Sorb or Carbo-Sorb E with Permafluor E<sup>+</sup>, the following conditions are recommended:

1. Carbo-Sorb E with Permafluor E<sup>+</sup>

For ratios of Carbo-Sorb E to Permafluor E<sup>+</sup> from 1:10 up to 1:1, maximum saturation of carbon dioxide is possible with no phase separation of the resulting carbamate.

## 2. Carbo-Sorb with Permafluor E<sup>+</sup>

When using Carbo-Sorb with Permafluor E<sup>+</sup> at any ratio up to 1:1 the recommendation is that 80% carbon dioxide capacity is not exceeded. Above this capacity there may be phase separation of the formed carbamate. If however this capacity is exceeded, then either of the following two methods will produce a clear homogeneous mixture:

- i) Add methanol to the mixture until a clear solution forms (usually 2-4 mL).
- ii) Add extra Carbo-Sorb to the mixture until a clear solution forms. This method simply dilutes the absorption capacity below the critical 80% level.

### Note:

The use of Carbo-Sorb or Carbo-Sorb E is not recommended in enzymatic, plant or human studies due to the corrosive nature of the volatile amine present.

### Summary

The information presented in the previous sections (1-6) of this publication are condensed into a quick reference table (Table 4). This may prove particularly useful when the total trapping capacity per standard 20 mL LSC vial is required.

CO <sub>2</sub> Absorber	mM CO <sub>2</sub> per mL	mL for 1 mM CO <sub>2</sub>	mL for 5mM CO <sub>2</sub>	mL for 10 mM CO <sub>2</sub>	LSC Cocktail	Cocktail Volume	mL of Absorber	Max. CO <sub>2</sub> Capacity (mM)
0.1 M Sodium & Potassium Hydroxide	0.05	20.00	-	-	Emulsifier-Safe Opti-Fluor	15.0 mL 14.0 mL	3.00 7.00	0.15 0.35
0.5 M Sodium & Potassium Hydroxide	0.25	4.00	-	-	Ultima-Flo AF	10.0 mL	10.00	2.50
1.0 M Sodium & Potassium Hydroxide	0.50 0.50	2.00 2.00	10.00 10.00	- -	Hionic-Fluor Ultima-Flo AF	14.0 mL 14.0 mL	7.00 7.00	3.50 3.50
1.0 M Hyamine Hydroxide in Methanol	0.50	2.00	10.00	-	Emulsifier-Safe Insta-Fluor	15.0 mL 12.0 mL	4.50 9.00	2.25 4.50
Ethanolamine	8.10	0.12	0.62	1.23	Hionic-Fluor + Methyl Cellosolve	10.0 mL 8.5 mL	3.00	24.30
Carbo-Sorb	5.80	0.17	0.86	1.72	Permafluor E+	10.0 mL	8.00	46.40

Carbo-Sorb E	4.80	0.21	1.04	2.08	Permafluor E+	10.0 mL	10.00	48.00
--------------	------	------	------	------	---------------	---------	-------	-------

**Table 4.**

Reference table for CO<sub>2</sub> trapping and LS counting.

## **Conclusion**

There are a variety of PerkinElmer LSC cocktails, both high flash-point and classical types, which are suitable for <sup>14</sup>C<sub>2</sub> absorption work regardless of the trapping agent used. If problems with trapping and counting persist, or an alternative trapping agent not covered in this publication is used, please call your local PerkinElmer representative for further applications support.

## **Recommended Literature**

1. Accurate Determination of <sup>14</sup>C<sub>2</sub> by Expulsion from Blood. Kaczmar, B.U. and Manet R. *Appl. Radiat. Isot.* Vol **38** No. 7, pp 577-578, 1987.
2. The Use of CO<sub>2</sub> Absorbers for the Determination of Specific <sup>14</sup>C Activities. R.M. Qureshi, Peter Fritz and R.J. Dsimmie, *Int. J. Appl. Radiat. Isot.* vol **36** No.2, pp 165-170, 1985.
3. <sup>14</sup>C Dating of Hydrological Samples Using Simple Procedures. Riffat M. Qureshi and Peter Fritz. *Int. J. Appl. Radiat. Isot.* vol **36** No. 10 p 825,1985.
4. A Device for the Liberation and Determination of <sup>14</sup>C<sub>2</sub>. Schadewaldt et. al. *Analytical Biochemistry* Vol **132**, pp 400-404, 1983.
5. A Method of Counting <sup>14</sup>C as CaCO<sub>3</sub> in a Liquid Scintillator with Improved Precision. Pfeiffer K., Rank, D. and Tschurlovits. *Int. J. of App. Radiat. Isot.* vol **32**, pp 665-667, 1981.
6. Improved Technique for Accurate and Convenient Assay of Biological Reactions Liberating <sup>14</sup>C<sub>2</sub>. Sissons, C.H. *Analytical Biochemistry* Vol **70**, pp 454-462, 1976.