

Caution: For Laboratory Use. A research reagent for research purposes only

AlphaLISA® Acceptor Beads

Product No.: 6772001, 6772002, 6772003, 6772004L

Lot No.: 660-068-A

Material Provided

Format:	6772001	1 mg
	6772003	50 mg
	6772002	5 mg
	6772004L	XXX mg (LQC)

Manufacturing Date: February 14, 2011

Kit Components:

Component	6772001 (1mg)	6772002 (5 mg)	6772003 (50 mg)	6772004L (LQC mg)
AlphaLISA Acceptor Beads at 20 mg/mL in 100 mM Tris, 0.05% Proclin-300, pH 8.0	1 x 50 µL (6772001B)	1 x 250 µL (6772002B)	5 x 500 µL (6772003B)	___ x ___ mL (6772004B)

Product Information

Stability: This product is stable for at least **24 months** from manufacturing date if used and stored under recommended conditions.

Storage Conditions: Store undiluted at 4°C protected from light. Freeze-thaw is not recommended and can cause the beads to form aggregates.

For additional information on running AlphaScreen® assays or on potential interfering compounds, please visit our website: www.perkinelmer.com/AlphaTech

Quality Control

Maximum signal and background counts were determined in an AlphaScreen® BSA-Biotin titration assay performed on an EnVision® HTS Alpha detection instrument. We certify that these results meet our requirements.

<u>TEST</u>	<u>RESULTS</u>
Maximum signal	67247 cps
Background	4746 cps

Suggested Materials and Instrumentation

Please visit our website

www.perkinelmer.com/AlphaTech

This product is not for resale or distribution except by authorized distributors.

LIMITED WARRANTY: PerkinElmer BioSignal Inc. warrants that, at the time of shipment, the products sold by it are free from defects in material and workmanship and conform to specifications which accompany the product. PerkinElmer BioSignal Inc. makes no other warranty, express or implied with respect to the products, including any warranty of merchantability or fitness for any particular purpose. Notification of any breach of warranty must be made within 60 days of receipt unless otherwise provided in writing by PerkinElmer BioSignal Inc. No claim shall be honored if the customer fails to notify PerkinElmer BioSignal Inc. within the period specified. The sole and exclusive remedy of the customer for any liability of PerkinElmer BioSignal Inc. of any kind including liability based upon warranty (express or implied whether contained herein or elsewhere), strict liability contract or otherwise is limited to the replacement of the goods or the refunds of the invoice price of goods. PerkinElmer BioSignal Inc. shall not in any case be liable for special, incidental or consequential damages of any kind.

PerkinElmer, Inc.
940 Winter Street
Waltham, MA 02451 USA
P: (800) 762-4000 or
(+1) 203-925-4602
www.perkinelmer.com



For a complete listing of our global offices, visit www.perkinelmer.com/ContactUs

Copyright ©2009, PerkinElmer, Inc. All rights reserved. PerkinElmer® is a registered trademark of PerkinElmer, Inc. All other trademarks are the property of their respective owners.